WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

5. Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

WelchAllyn®

Spot Vital Signs LXi Operating Instructions

On/Off Button
Controls power to the device.

Select Button
Accepts the current option.

Blood Pressure Operation

Press the **BP Start/Stop** button to start a blood pressure measurement; press it again to stop a current reading.

SureTemp Plus Temperature Operation

1. Remove probe from well and apply a new probe cover. Use the **Navigation** buttons to change the temperature mode then press the **Select** button.
2. Place probe in contact with patient tissue as directed in the Directions for Use manual. When done, a beep sounds. The temperature appears on the display.
3. Eject probe cover. Return probe to well before starting another measurement.

Braun ThermoScan® PRO 4000 Temperature Operation

1. Remove probe from holder and apply a new probe cover; when ready, the device will beep.
2. Fit probe snugly in the ear canal as directed in the Directions for Use manual.
3. Press the Activation button. When done, a long beep sounds. The probe will display the temperature.
4. Eject probe cover and return probe to the holder. The Spot Vital Signs LXi will display the temperature.

SpO_2 Operation

Apply sensor to patient. Verify there is a signal on the pulse signal bar graph. The SpO_2 measurement and pulse rate will update continuously for 10 minutes.

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

705298 Rev. C

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

705298 Rev. C

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

705298 Rev. C

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

705298 Rev. C

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

705298 Rev. C

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

705298 Rev. C

Save/Send Operation

1. Use the **Navigation** buttons to move the arrow to “Save/Next Reading” or “Send/Next Reading” in the Navigation window.
2. Press the **Select** button. Spot Vital Signs LXi will transfer data electronically if it can or will save data in memory if it cannot transfer it. A blank status window will appear.

Manual Weight, Height, Respiration, Pain Entry

You may enter these parameters if they are enabled in the Internal Configuration Mode; see Directions for Use manual for details.

1. Press the **Navigation** buttons to move the arrow in front of the desired parameter in the Navigation window.
2. Press the **Select** button.
3. Press the **Navigation** buttons up to increase values or down to decrease values.
4. Press the **Select** button to accept the entry.

Memory/Print/Erase Operation

The Memory button allows you to scroll through, print, or erase readings.

MEM

1. Press the **Memory** button. Spot Vital Signs LXi will display the previous reading.
2. Press the **Navigation** buttons to select the desired function and press the **Select** button. Follow instructions on the display.

Caution: Consult accompanying documents.

705298 Rev. C